

Office: 802-540-0148 | Fax: 802-540-0147 480 HERCULES DR. COLCHESTER, VT 05446

Company: Kria Commons 8 Harbor View Rd Burlington, VT 05403 Customer ID: 190904-01 Grower License #: N/A

Certificate of Analysis

Sample ID: CO2 THC Crude (MIX1) Lot: MANU0005-T-NCC-THCMIX1-C1-3 Matrix: Distillate Date Sampled: 12/20/2022 Date Received: 12/22/2022

Report Date: 1/10/2023 Date Analyzed: 1/6/2023 Analyst: 042 Report ID: C221222AG

Heavy Metal Summary

Heavy Metal Profile	LOQ (ppm)	Concentration (ppm)
Arsenic (As)	0.0001	0.0010
Cadmium (Cd)	0.0001	<loq< th=""></loq<>
Mercury (Hg)	0.0001	<loq< th=""></loq<>
Lead (Pb)	0.0001	0.0030



N/A
Percent Moisture

Heavy Metal Methodology: ICP-MS using PerkinElmer $\texttt{NexION}^{\circledast}$ 2000 ICP Mass Spectrometer

Reagent Blanks: < LOQs for all analytes

ppm = parts per million

LOQ = The lowest quantity that this method can reliably detect. Any heavy metal that was not detected is assumed to be less than the stated LOQ (<LOQ).

All results reflect dry weight of material, based on % moisture of the sample.

All moisture analysis is determined by loss-on-drying measurement using OHAUS Model MB90 Moisture Content Readers.

This report shall not be reproduced except in full without approval of the laboratory. This is to provide assurance that parts of a report are not taken out of context. Results apply to the *Certified by:* samples as received.

Luke E.M.

Luke Emerson Mason (Laboratory Director, Bia Diagnostics)